

AS and A-level Chemistry practicals: Equipment set up

Practical 9: Investigate how pH changes when a weak acid reacts with a strong base and when a strong acid reacts with a weak base

pH changes

- There are currently no photographs for this practical.
- pH meters should be calibrated before use in accordance with the manufacturers instructions and allowed to stabilise in distilled/deionised water for few hours before the practical to ensure the probes are working correctly.